

Research article

Open Access

## Comparison of the MicroScan, VITEK 2, and Crystal GP with 16S rRNA sequencing and MicroSeq 500 v2.0 analysis for coagulase-negative Staphylococci

Miyoung Kim<sup>†1</sup>, Se Ran Heo<sup>2</sup>, Soon Hee Choi<sup>2</sup>, Hyelin Kwon<sup>2</sup>, Jeong Su Park<sup>1</sup>, Moon-Woo Seong<sup>3</sup>, Do-Hoon Lee<sup>3</sup>, Kyoung Un Park<sup>\*1,2,4</sup>, Junghan Song<sup>1,2,4</sup> and Eui-Chong Kim<sup>1,4</sup>

Address: <sup>1</sup>Department of Laboratory Medicine, Seoul National University Hospital, 101 Daehang-no, Jongno-gu, Seoul, South Korea, <sup>2</sup>Department of Laboratory Medicine, Seoul National University Bundang Hospital, 300 Gumi-dong, Bundang-gu, Seongnam-si, Gyeonggi-do, South Korea, <sup>3</sup>Department of Laboratory Medicine, National Cancer Center, 809 Madu-1-dong, Ilsandong-gu, Goyang, Gyeonggi-do, South Korea and <sup>4</sup>Department of Laboratory Medicine, Seoul National University College of Medicine, 101 Daehang-no, Jongno-gu, Seoul, South Korea

Email: Miyoung Kim - rabbit790622@gmail.com; Se Ran Heo - sellyozoo@hanmail.net; Soon Hee Choi - soon2ya@snu.ac.kr; Hyelin Kwon - hyelinn@hanmail.net; Jeong Su Park - mdjs0721@naver.com; Moon-Woo Seong - mwseong@ncc.re.kr; Do-Hoon Lee - dhlee@ncc.re.kr; Kyoung Un Park\* - m91w95@dreamwiz.com; Junghan Song - songjhcp@snu.ac.kr; Eui-Chong Kim - euichong@snu.ac.kr

\* Corresponding author †Equal contributors

Published: 23 December 2008

Received: 10 June 2008

BMC Microbiology 2008, 8:233 doi:10.1186/1471-2180-8-233

Accepted: 23 December 2008

This article is available from: <http://www.biomedcentral.com/1471-2180/8/233>

© 2008 Kim et al; licensee BioMed Central Ltd.

This is an Open Access article distributed under the terms of the Creative Commons Attribution License (<http://creativecommons.org/licenses/by/2.0>), which permits unrestricted use, distribution, and reproduction in any medium, provided the original work is properly cited.

### Abstract

**Background:** Three phenotypic identification systems (MicroScan, VITEK 2, and Crystal GP) were evaluated for their accuracy to identify coagulase-negative staphylococci (CNS). A total of 120 clinical isolates confirmed to be CNS via 16S rRNA sequencing and analysis with the MicroSeq 500 v2.0 database were assessed.

**Results:** The MicroScan, VITEK 2, and Crystal GP systems correctly identified 82.5%, 87.5%, and 67.5% of the isolates, respectively. Misidentification was the main problem in MicroScan (10.8%) and Crystal GP (23.3%) systems, whereas the main problem of VITEK 2 was low-level discrimination (7.5%).

**Conclusion:** None of the 3 phenotypic systems tested could accurately and reliably identify CNS at the species level. Further verifications such as biochemical testing or 16S rRNA sequencing together with analysis using a comparable database might be helpful in this regard.

### Background

Because of their ubiquity and low virulence, coagulase-negative staphylococci (CNS) have generally been considered to be nonpathogens or simple contaminants. Recently, their clinical significance is being increasingly recognized with the elucidation of their pathogenicity. CNS can form biofilms and have been demonstrated to

exhibit antibiotic resistance [1,2]. *S. epidermidis*, *S. haemolyticus*, and *S. lugdunensis*, to a lesser extent, are well-known etiological agents of implanted device-mediated infections [2-5]. *S. saprophyticus* can cause community-acquired infections of the uropoietic tract. Hence, species-level identification of CNS is necessary for correct guidance of clinicians in terms of appropriate treatment strat-

egy, and a wide variety of identification methods have been proposed.

Many automated phenotypic identification systems are commercially available, including the MicroScan (Dade Behring, West Sacramento, CA, USA), VITEK 2 (bioMérieux, Maray l'Etoile, France), and Crystal GP (Becton Dickinson, Sparks, MD, USA) systems. On the basis of metabolic activities and/or morphological features, these systems enable microbiologists to identify bacterial isolates at the species level with greater ease, accuracy, and rapidity than that previously achieved [6]. However, these systems have several potential problems: (i) different strains in one species may not exhibit a specific characteristic, (ii) isolates from old cultures may not show the expected biochemical patterns, (iii) isolates from a host who has undergone long-term antimicrobial therapy may alter their typical biochemical characteristics, (iv) the same strain may not yield the same results in repeated tests, (v) the databases have data on a limited number of species, (vi) phenotypic variation may affect the accuracy of species-level identification by automated phenotypic systems, and (vii) phenotypic systems often suggest 2 or more designations with comparable probability levels [6-9].

Recently, genotypic methods are emerging as the new standard for bacterial identification in automated laboratories [4,6,7,10-14]. 16S ribosomal RNA (rRNA) sequences comprising both variable and conserved regions allow for clear differentiation between organisms not only at the species level but also at the subspecies level [10]. These sequences permit better identification of rare or phenotypically aberrant species as well as nonculturable bacteria. The MicroSeq 500 system (Applied Biosystems Inc. [ABI], Foster City, CA, USA) is a commercially available software system for 16S rRNA analysis.

Despite the increasing clinical significance of CNS and the growing use of automated phenotypic systems in clinical laboratories, few studies have systematically evaluated the accuracy of these systems [4,9,15]. In the present study, the accuracy of 3 commercial phenotypic systems (MicroScan, VITEK 2, and Crystal GP) for CNS identification was compared. 16S rRNA sequencing and MicroSeq 500 analysis were used as references [10-12,16-19].

## Results

### Clinical isolates

By using in-house primers and the MicroSeq 500 v2.0 database, we successfully identified all the 120 clinical isolates with more than 97% matches (data not shown). The identified species were as follows: 16 *S. capitis*, 4 *S. caprae*, 3 *S. cohnii*, 53 *S. epidermidis*, 4 *S. haemolyticus*, 25 *S. hominis*, 6 *S. lugdunensis*, 1 *S. saprophyticus*, 2 *S. simulans*,

and 6 *S. warneri*. Blood culture specimens primarily contained *S. epidermidis*, whereas a variety of uncommon CNS were isolated from pitted keratolysis specimens. The result reflected the typical distribution of staphylococcal species detected routinely in microbiology laboratories.

### Identification using MicroScan

The overall performances of 3 phenotypic systems are summarized in Table 1 and the incorrect identification results are listed in Table 2. Of 120 isolates, correct identification, low-level discrimination and misidentification were 82.5% (99), 6.7% (8) and 10.8% (13), respectively. Nonidentification was not observed. Three *S. caprae* isolates were not correctly identified (isolate number: M211, M215, and M225), because the species *S. caprae* is not included in the MicroScan version 2.0 or 2.1 database. One of the *S. epidermidis* isolates (M106) was misidentified as *S. aureus*.

### Identification using VITEK 2

In the VITEK 2 analysis, correct identification, low-level discrimination and misidentification were 87.5% (105), 7.5% (9) and 5.0% (6), respectively. Nonidentification did not occur. Of note, the analysis of 6 *S. epidermidis* isolates (6/53, 11.3%) resulted in low-level discriminations – multiple identifications as *S. hominis*, *S. epidermidis*, or *Aerococcus viridans* with low percent probabilities, suggesting *S. hominis* with the highest percent probability followed by *S. epidermidis* (M123, M124, M125, M126, M127, M137).

### Identification using Crystal GP

Crystal GP is not as fully automated as the other two, as the reading of the biochemical reaction depends on manual reading. In order to prevent the bias resulting from manual reading, the biochemical reaction results were read by two different technicians, followed by repeated tests on wrongly identified isolates or those with multiple identifications or nonidentifications. The identification results of 81 isolates (67.5%) using the Crystal GP system agreed with those of the MicroSeq 500 system with confidence factors of more than 0.9 (correct identifications). In total, 23.3% (28) were misidentified, and 9.2% (11) could not be identified (nonidentification). Low-level discrimination was not observed. Misidentifications at the genus level occurred for 1 *S. epidermidis* and 1 *S. hominis* isolate (M112, M216). The Crystal GP system correctly identified all the isolates of *S. cohnii*, *S. haemolyticus*, *S. saprophyticus*, *S. simulans*, and *S. warneri*.

## Discussion

With the increased attention being given to the clinical significance of CNS, clinical laboratories must correctly identify CNS at the species level by using reliable and reproducible methods. Because clinical microbiology lab-

**Table 1: Clinical isolates and identification results from MicroScan, VITEK2, and Crystal GP**

Clinical Isolates (using MicroSeq 500)	No of isolates	Correct Identification <sup>a</sup>	Low-level Discrimination <sup>b</sup>	Misidentification <sup>c</sup>	Nonidentification <sup>d</sup>
MicroScan					
<i>S. capitis</i>	16	13 (81.3%)	-	3 (18.8%)	-
<i>S. caprae</i>	4	-	-	4 (100.0%)	-
<i>S. cohnii</i>	3	1 (33.3%)	1 (33.3%)	1 (33.3%)	-
<i>S. epidermidis</i>	53	51 (96.2%)	-	2 (3.8%)	-
<i>S. haemolyticus</i>	4	3 (75.0%)	1 (25.0%)	-	-
<i>S. hominis</i>	25	23 (92.0%)	2 (8.0%)	-	-
<i>S. lugdunensis</i>	6	2 (33.3%)	1 (16.7%)	3 (50.0%)	-
<i>S. saprophyticus</i>	1	1 (100.0%)	-	-	-
<i>S. simulans</i>	2	2 (100.0%)	-	-	-
<i>S. warneri</i>	6	3 (50.0%)	3 (50.0%)	-	-
Subtotal	120	99 (82.5%)	8 (6.7%)	13 (10.8%)	-
VITEK 2					
<i>S. capitis</i>	16	16 (100.0%)	-	-	-
<i>S. caprae</i>	4	3 (75.0%)	-	1 (25.0%)	-
<i>S. cohnii</i>	3	3 (100.0%)	-	-	-
<i>S. epidermidis</i>	53	47 (88.7%)	6 (11.3%)	-	-
<i>S. haemolyticus</i>	4	4 (100.0%)	-	-	-
<i>S. hominis</i>	25	22 (88.0%)	1 (4.0%)	2 (8.0%)	-
<i>S. lugdunensis</i>	6	5 (83.3%)	1 (16.7%)	-	-
<i>S. saprophyticus</i>	1	1 (100.0%)	-	-	-
<i>S. simulans</i>	2	1 (50.0%)	1 (50.0%)	-	-
<i>S. warneri</i>	6	3 (50.0%)	-	3 (50.0%)	-
Subtotal	120	105 (87.5%)	9 (7.5%)	6 (5.0%)	-
Crystal GP					
<i>S. capitis</i>	16	13 (81.3%)	-	3 (18.8%)	-
<i>S. caprae</i>	4	-	-	4 (100.0%)	-
<i>S. cohnii</i>	3	3 (100.0%)	-	-	-
<i>S. epidermidis</i>	53	49 (92.5%)	-	3 (5.7%)	1 (1.9%)
<i>S. haemolyticus</i>	4	4 (100.0%)	-	-	-
<i>S. hominis</i>	25	1 (4.0%)	-	15 (60.0%)	9 (36.0%)
<i>S. lugdunensis</i>	6	2 (33.3%)	-	3 (50.0%)	1 (16.7%)
<i>S. saprophyticus</i>	1	1 (100.0%)	-	-	-
<i>S. simulans</i>	2	2 (100.0%)	-	-	-
<i>S. warneri</i>	6	6 (100.0%)	-	-	-
Subtotal	120	81 (67.5%)	-	28 (23.3%)	11 (9.2%)

Correct identification<sup>a</sup>: single, unambiguous, correct identification at the species level

Low-level discrimination<sup>b</sup>: two or more possible species level identification including the correct one

Misidentification<sup>c</sup>: genus or species-level identification different from that obtained with the reference method

Nonidentification<sup>d</sup>: no identification or unidentification

oratories are increasingly becoming dependent on automated systems, the accuracy of these systems must be investigated. In the present study, the accuracy of 3 phenotypic identification systems for CNS identification was tested and compared with the sequencing of 16S rRNA.

MicroScan and VITEK 2 showed similar concordance rate to MicroSeq 500. VITEK 2 showed a slightly higher "correct identification" rate (82.5% vs. 87.5%, respectively). When the low-level discrimination results that included

the correct identification were considered together, the concordance rates were 89.2% (107/120) for MicroScan and 95.0% (114/120) for VITEK 2. Crystal GP had the lowest correct identification rate (67.5%). All 3 systems had misidentifications: single but incorrect identification results with higher than acceptable probability of over 85%, which is consistent with the system manufacturer's reported homology value without additional verification, or with higher than a confidence factor of 0.9. This indicates that CNS identifications made using these pheno-

**Table 2: Identification results with percent probability or confidence factor of single wrong identification or multiple identifications in three phenotypic systems**

	MicroSeq 500	MicroScan	VITEK2	Cystal GP
M102	<i>S. epidermidis</i>	<i>S. epidermidis</i> 91.36%	<i>S. epidermidis</i> 97.10%	<i>S. intermedius</i>
M104	<i>S. epidermidis</i>	<i>S. xylosum</i> 95.12%	<i>S. epidermidis</i> 99.00%	<i>S. epidermidis</i>
M106	<i>S. epidermidis</i>	<i>S. aureus</i> 87.13%	<i>S. epidermidis</i> 99.00%	<i>S. epidermidis</i>
M107	<i>S. capitis</i>	<i>S. simulans</i> 88.95%	<i>S. capitis</i> 98.78%	<i>S. capitis</i>
M112	<i>S. epidermidis</i>	<i>S. epidermidis</i> 99.99%	<i>S. epidermidis</i> 99.00%	<i>K. sedentarius</i>
M117	<i>S. lugdunensis</i>	<i>S. haemolyticus</i> 90.39%	<i>S. lugdunensis</i> 99.00%	Nonidentification
M118	<i>S. lugdunensis</i>	<i>S. haemolyticus</i> 96.86%	<i>S. lugdunensis</i> 90.37%	<i>S. haemolyticus</i>
M120	<i>S. hominis</i>	<i>S. hominis</i> 94.57%	<i>S. hominis</i> 95.00%	Nonidentification
M123	<i>S. epidermidis</i>	<i>S. epidermidis</i> 98.20%	<i>S. hominis</i> , <i>S. epidermidis</i> , <i>A. viridans</i> 33.82%, 33.09%, 33.09%	<i>S. epidermidis</i>
M124	<i>S. epidermidis</i>	<i>S. epidermidis</i> 99.25%	<i>S. hominis</i> , <i>S. epidermidis</i> 50.27%, 49.73%	Nonidentification
M125	<i>S. epidermidis</i>	<i>S. epidermidis</i> 98.20%	<i>S. hominis</i> , <i>S. epidermidis</i> 50.27%, 49.73%	<i>S. epidermidis</i>
M126	<i>S. epidermidis</i>	<i>S. epidermidis</i> 99.99%	<i>S. hominis</i> , <i>S. epidermidis</i> , <i>A. viridans</i> 33.71%, 33.33%, 32.96%	<i>S. epidermidis</i>
M127	<i>S. epidermidis</i>	<i>S. epidermidis</i> 99.99%	<i>S. hominis</i> , <i>S. epidermidis</i> 50.28%, 47.92%	<i>S. epidermidis</i>
M128	<i>S. hominis</i>	<i>S. hominis</i> 94.58%	<i>S. hominis</i> 94.86%	<i>S. saprophyticus</i>
M130	<i>S. capitis</i>	<i>S. epidermidis</i> 90.20%	<i>S. capitis</i> 92.00%	<i>S. capitis</i>
M137	<i>S. epidermidis</i>	<i>S. epidermidis</i> 95.28%	<i>S. hominis</i> , <i>S. epidermidis</i> 50.56%, 49.44%	<i>S. epidermidis</i>
M138	<i>S. capitis</i>	<i>S. capitis</i> 95.28%	<i>S. capitis</i> 98.86%	<i>S. auricularis</i>
M139	<i>S. capitis</i>	<i>S. capitis</i> 99.60%	<i>S. capitis</i> 98.86%	<i>S. hominis</i>
M141	<i>S. epidermidis</i>	<i>S. epidermidis</i> 99.99%	<i>S. epidermidis</i> 98.16%	<i>S. hominis</i>
M149	<i>S. capitis</i>	<i>S. epidermidis</i> 90.20%	<i>S. capitis</i> 92.00%	<i>S. capitis</i>
M153	<i>S. hominis</i>	<i>S. hominis</i> 97.44%	<i>S. hominis</i> 99.00%	Nonidentification
M154	<i>S. hominis</i>	<i>S. hominis</i> 94.62%	<i>S. hominis</i> 98.35%	Nonidentification
M155	<i>S. hominis</i>	<i>S. hominis</i> 97.44%	<i>S. hominis</i> 97.21%	<i>S. warneri</i> , <i>S. saprophyticus</i> 0.5574, 0.4415
M156	<i>S. hominis</i>	<i>S. hominis</i> 95.00%	<i>S. hominis</i> 97.00%	<i>S. saprophyticus</i>
M157	<i>S. warneri</i>	<i>S. warneri</i> , <i>S. cohnii</i> , <i>S. capitis</i> , etc. 60.92%, 15.78%, 12.67%	<i>S. cohnii</i> , <i>S. vitulinus</i> 50.54%, 49.46%	<i>S. warneri</i>
M158	<i>S. warneri</i>	<i>S. warneri</i> 90.92%	<i>S. cohnii</i> 90.54%	<i>S. warneri</i>
M160	<i>S. hominis</i>	<i>S. hominis</i> , <i>S. capitis</i> , <i>S. warneri</i> , etc. 37.58%, 24.33%, 20.21%	<i>S. hominis</i> 93.79%	Nonidentification
M161	<i>S. hominis</i>	<i>S. hominis</i> 94.57%	<i>S. hominis</i> 96.98%	<i>S. saprophyticus</i>
M162	<i>S. caprae</i>	<i>S. hominis</i> 92.35%	<i>S. auricularis</i> 90.40%	<i>S. saprophyticus</i>
M163	<i>S. hominis</i>	<i>S. hominis</i> 98.82%	<i>S. hominis</i> 98.50%	Nonidentification
M166	<i>S. hominis</i>	<i>S. hominis</i> 85.00%	<i>S. hominis</i> 98.82%	Nonidentification
M167	<i>S. hominis</i>	<i>S. hominis</i> 93.05%	<i>S. caprae</i> 96.97%	<i>S. caprae</i>
M176	<i>S. simulans</i>	<i>S. simulans</i> 96.77%	<i>S. simulans</i> , <i>S. warneri</i> 50.28%, 49.72%	<i>S. simulans</i>
M17	<i>S. hominis</i>	<i>S. simulans</i> 99.09%	<i>S. haemolyticus</i> 96.74%	<i>S. haemolyticus</i>
M179	<i>S. hominis</i>	<i>S. epidermidis</i> , <i>S. hominis</i> 68.34%, 31.66%	<i>S. hominis</i> 94.88%	Nonidentification
M180	<i>S. hominis</i>	<i>S. hominis</i> 92.77%	<i>S. hominis</i> 93.56%	<i>S. warneri</i>
M181	<i>S. hominis</i>	<i>S. hominis</i> 92.77%	<i>S. hominis</i> , <i>S. warneri</i> 50.27%, 49.73%	<i>S. warneri</i>
M182	<i>S. hominis</i>	<i>S. hominis</i> 94.57%	<i>S. hominis</i> 95.00%	Nonidentification
M183	<i>S. hominis</i>	<i>S. hominis</i> 88.68%	<i>S. hominis</i> 91.00%	<i>S. saprophyticus</i>
M184	<i>S. hominis</i>	<i>S. hominis</i> 92.13%	<i>S. hominis</i> 99.00%	<i>S. intermedius</i>
M185	<i>S. hominis</i>	<i>S. hominis</i> 96.01%	<i>S. hominis</i> 96.81%	<i>S. capitis</i>
M186	<i>S. hominis</i>	<i>S. hominis</i> 98.50%	<i>S. hominis</i> 95.00%	<i>S. aureus</i> , <i>S. intermedius</i>
M187	<i>S. hominis</i>	<i>S. hominis</i> 98.50%	<i>S. hominis</i> 99.00%	Nonidentification
M203	<i>S. hominis</i>	<i>S. hominis</i> 97.44%	<i>S. hominis</i> 97.27%	<i>S. cohnii</i>
M205	<i>S. warneri</i>	<i>S. warneri</i> , <i>S. capitis</i> , <i>S. hominis</i> 86.37%, 7.96%, 5.67%	<i>S. saprophyticus</i> 98.03%	<i>S. warneri</i>
M207	<i>S. lugdunensis</i>	<i>S. haemolyticus</i> 98.08%	<i>S. lugdunensis</i> 99.00%	<i>S. lugdunensis</i>
M208	<i>S. haemolyticus</i>	<i>S. haemolyticus</i> , <i>S. simulans</i> , <i>S. warneri</i> 54.38%, 41.97%, 3.66%	<i>S. haemolyticus</i> 94.74%	<i>S. haemolyticus</i>
M209	<i>S. hominis</i>	<i>S. hominis</i> 97.44%	<i>S. hominis</i> 97.76%	<i>S. cohnii</i>
M210	<i>S. cohnii</i>	<i>S. cohnii</i> , <i>S. xylosum</i> 67.28%, 29.12%	<i>S. cohnii</i> 99.00%	<i>S. cohnii</i>
M211	<i>S. caprae</i>	<i>S. aureus</i> , <i>S. capitis</i> 69.87%, 29.25%	<i>S. caprae</i> 98.60%	<i>S. intermedius</i>
M212	<i>S. capitis</i>	<i>S. capitis</i> 99.11%	<i>S. capitis</i> 98.56%	<i>S. vitulus</i>
M213	<i>S. warneri</i>	<i>S. auricularis</i> , <i>S. hominis</i> , <i>S. warneri</i> 45.13%, 27.55%, 24.55%	<i>S. warneri</i> 95.00%	<i>S. warneri</i>

**Table 2: Identification results with percent probability or confidence factor of single wrong identification or multiple identifications in three phenotypic systems (Continued)**

M214	<i>S. lugdunensis</i>	<i>S. lugdunensis</i> 98.81%	<i>S. lugdunensis</i> 95.00%	<i>S. simulans</i>
M215	<i>S. caprae</i>	<i>S. capitis</i> , <i>S. aureus</i> 62.82%, 35.76%	<i>S. caprae</i> 99.00%	<i>S. cohnii</i>
M216	<i>S. hominis</i>	<i>S. hominis</i> 97.44%	<i>S. hominis</i> 97.76%	<i>M. kristinae</i>
M221	<i>S. lugdunensis</i>	<i>S. lugdunensis</i> , <i>S. haemolyticus</i> , <i>S. hominis</i> 62.62%, 28.34%, 9.04%	<i>S. lugdunensis</i> , <i>S. warneri</i> 50.00%, 50.00%	<i>S. lugdunensis</i>
M222	<i>S. lugdunensis</i>	<i>S. lugdunensis</i> 96.30%	<i>S. lugdunensis</i> 91.26%	<i>S. simulans</i>
M224	<i>S. cohnii</i>	<i>S. xylosum</i> 91.70%	<i>S. cohnii</i> 99.00%	<i>S. cohnii</i>
M225	<i>S. caprae</i>	<i>S. warneri</i> , <i>S. simulans</i> , <i>S. epidermidis</i> 70.19%, 13.84%, 12.32%	<i>S. caprae</i> 99.00%	<i>S. capitis</i>

typic systems cannot always be considered accurate, regardless of the reported probability.

In MicroScan analysis, none of the *S. caprae* identifications were correct (3 low-level identifications and 1 misidentification). This is because the species *S. caprae* is not included in the MicroScan database, which illustrates the inappropriateness of this database. Ambiguous identifications, including low-level discriminations and misidentifications, were observed for 8 out of 10 species (with the exception of *S. saprophyticus* and *S. warneri*) with the MicroScan system, compared with 6 ambiguous identifications with the VITEK2 system and 5 with the Crystal GP system. This finding was consistent with results reported by other researchers who found that the MicroScan system has greater accuracy in the identification of *S. epidermidis* and *S. saprophyticus* than for identification of other species [20-22]. The other notable finding of the present study is that 1 *S. epidermidis* isolate from blood culture was misidentified as *S. aureus* with 87.5% probability, which could have delayed appropriate treatment for the patient.

In VITEK 2 analysis, low-level discriminations of 6 *S. epidermidis* isolates (11.3% of *S. epidermidis* isolates) occurred, compared to higher correct identification rates with the MicroScan and Crystal GP systems (96.2% and 92.5%, respectively). Some researchers reported different sort of difficulties in identifying *S. epidermidis* with VITEK 2 [9,23,24]. However, others reported high accuracy of VITEK 2 for the identification of staphylococcal species [25,26]. The performance of VITEK 2 in the identification of CNS including *S. epidermidis* seems not to be determined yet, and the additional testing such as manual biochemical testing or sequencing is required for the accurate diagnosis.

The Crystal GP system provides no guidelines for the interpretation of the confidence factor, therefore, all isolates with confidence factors lower than 0.9 were retested. In spite of repeated testing, this system had the lowest correct identification rate and the highest rate of nonidentification.

One limitation of this study is that additional biochemical testing was not performed. 16S rRNA sequencing might not be a perfect method for interspecies discrimination. However, not all the suggested biochemical tests are available in many automated laboratories or based on our experience, manual testing could not resolve the ambiguity of CNS identification in many cases. To improve reliability in this study, all isolates with discrepancies or uncertain results were retested using isolates preserved in skim milk at -70°C.

## Conclusion

None of the tested systems (MicroScan, VITEK 2, Crystal GP) are able to accurately identify all of the staphylococcal species evaluated in this study. Additional verification via molecular testing or suggested biochemical tests may prove helpful for correct CNS identification.

## Methods

### Clinical isolates

A total of 120 clinical isolates (95 blood culture isolates, 25 skin isolates from pitted keratolysis patients) were used in this study [27]. To exclude simple contaminants, only isolates of blood cultures that grew in more than 2 of 3 blood cultures were included. The pitted keratolysis specimen was obtained after repeated washing of the lesion with normal saline by scratching the lesion with a sterile blade into a sterile conical tube. The isolates were tested within 24 hours. For additional tests or repeated tests, the isolates were suspended in skim milk, stored at -70°C and retested. All of the isolates were confirmed to be CNS by use of 16S rRNA sequencing with MicroSeq 500 software (version 2.0) with over 97% matches.

### Phenotypic identification of CNS

Three phenotypic systems were evaluated: MicroScan Pos Combo panel type 1A (Dade Behring, West Sacramento, CA, USA), VITEK 2 GP-ID cards (bioMérieux, Marcy l'Etoile, France), and Crystal Gram Positive ID systems (Becton Dickinson, Sparks, MD, USA). Bacterial suspensions were prepared from well-isolated colonies. The colonies were suspended in the broth provided by each company, and the turbidity was adjusted to 0.5 McFarland. MicroScan Pos Combo panel type 1A was inocu-

lated according to the manufacturer's recommendations and processed with a MicroScan Walkaway 96 apparatus. The blood culture isolates were analyzed using a MicroScan LabPro system (version 2.0) and the skin isolates were analyzed with version 2.01. VITEK 2 GP-ID cards were used in combination with a VITEK 2 system (version 4.02) for all isolates. The Crystal GP results were read manually by 2 different technicians using the Crystal GP Identification Color Chart.

The MicroScan system classifies identification results into 4 categories by percent probability: species identified with high probability (> 85%), species identified with low probability (< 85%), very rare biotype, and group ID (> 85% but genus or group identified only). The VITEK 2 system categorizes the identification results by probability: excellent (96%–99%), very good (93%–95%), good (89%–92%), acceptable (85%–88%), low discrimination (2–3 identifications with 100% probability in total), and nonidentification. The Crystal GP system gives confidence factors instead of probability; however, there are no such criteria in terms of the confidence factor. When the confidence factor is low, Crystal GP reports non-identification, which necessitates additional manual testing.

The results of incorrect identifications (not concordant to that of MicroSeq 500) regardless of probability, identification with unacceptable probability (lower than 85% for MicroScan and VITEK 2) or low confidence factor (lower than 0.9 for Crystal GP), or nonidentification were retested with colonies that were stored in skim milk at -70°C. Those with incorrect, or consistently unacceptable or nonidentifiable results are listed in Table 2. The confidence factor of the Crystal GP system is not included, as there are no guidelines for these interpretations.

### **16S rRNA sequencing**

To extract bacterial DNA, InstaGene Matrix (Bio-Rad Laboratories, Hercules, CA, USA) was used. Two or 3 colonies of each isolate were suspended in 100 µl sterile distilled water and centrifuged for 5 min at 15,000 rpm. A 200 µl aliquot of InstaGene Matrix (Bio-Rad Laboratories, Hercules, CA, USA) was added to the sediment. The mixture was heat-lysed for 5 min at 100°C, cooled at room temperature, and centrifuged for 3 min at 15,000 rpm. The supernatant (2.5 µl) of each bacterial extract was used for successive amplification procedures. Polymerase chain reaction (PCR) was conducted in a total volume of 25 µl containing 2.5 mM dNTP, 10 pmol of each PCR primer, 0.6 units Taq polymerase, 2.5 µl 10 × PCR buffer with 15 mM MgCl<sub>2</sub> (Takara Bio, Inc., Shiga, Japan), and 2.5 µl template. In-house primers were designed using the Light-Cycler Probe Design software (version 2.0) (Roche, Penzberg, Germany), using published studies as a reference [8,9]. The primers used for amplification were MSQ-F (5'-

TGA AGA GTT TGA TCA TGG CTC AG-3') and MSQ-R (5'-ACC GCG GCT GCT GGC AC-3'). The PCR conditions were as follows: 10 min of initial denaturation at 95°C, followed by 35 cycles of annealing of 30s at 95°C, 30s at 60°C, 45s at 72°C, and a final 10-minute extension at 72°C. Gel electrophoresis was used to detect positive PCR signals and to confirm the length of the amplicon. Prior to sequencing, the PCR products were purified using ExoSAP-IT reagent (USB Corporation, Cleveland, Ohio, USA) according to the manufacturer's instructions. Forward and reverse sequencing reactions were conducted for each of the amplified products. The sequencing reaction consisted of 10 µl MicroSeq 500 sequencing mix (containing 1.6 pmol of MSQ-F and MSQ-R primers), 2.9 µl sterile distilled water, and 1 µl purified PCR product. Sequencing reactions were performed using Big Dye terminator reagents on an ABI Prism 3130xl Genetic Analyzer (Applied Biosystems Inc., Foster City, CA, USA) according to standard automated sequencer protocols.

### **Sequence analysis**

MicroSeq 500 version 2.0 software was used for sequence analysis. The analysis steps were as follows: (i) the forward and reverse sequences were assembled into one consensus sequence, (ii) the consensus sequence was edited to resolve discrepancies between the 2 strands by evaluating the electropherograms, and (iii) the consensus sequence was compared with sequences in the MicroSeq 500 database. The comparison, using the full alignment tool of the MicroSeq 500 software, generated a list of the closest matches with a distance score, which indicates the percentage difference between the unknown sequence and the database sequence. Only when the consensus sequence had more than a 97% match with that of the MicroSeq 500 results, the identification results were considered to be acceptable [7,10].

### **Analysis of the results**

The results of 16S rRNA sequencing and MicroSeq 500 analysis were regarded as the definitive identification [10-12,16-19]. The species-level identification and percent probability from phenotypic systems were all taken into consideration. The concordance of the results of the phenotypic systems with those of MicroSeq 500 were classified into 4 categories according to the following definitions: (i) correct identification (unambiguous, single identification identical to that of MicroSeq 500 at the species level, with > 85% probability in MicroScan and VITEK 2 or 0.9 confidence factor); (ii) low-level discrimination (2 or more possible species-level identifications, including the correct one, with probability < 85% in MicroScan and VITEK 2 or the low confidence factor under 0.9 in Crystal GP); (iii) misidentification (genus-level or species-level identification different from that obtained with the reference method) and (iv) nonidenti-

fication (isolates that were unable to be identified at the species level in spite of the repeated testing).

### Authors' contributions

MK designed the study, helped with the phenotypic identification and molecular studies, and wrote the manuscript. SHC performed sample collection and phenotypic identification. SRH, HK and JSP carried out the molecular genetic studies, participated in the sequence alignment. MWS, DHL, JS and ECK helped design the study and helped with technical issues. KUP conceived of study, and participated in its design and coordination, helped to draft the manuscript and reviewed the manuscript. All of the authors have read and approved the final manuscript.

### References

- von Eiff C, Peters G, Heilmann C: **Pathogenesis of infections due to coagulase-negative staphylococci.** *Lancet Infectious Dis* 2000, **2**:677-685.
- Longauerova A: **Coagulase negative staphylococci and their participation in pathogenesis of human infections.** *Bratisl Lek Listy* 2006, **107**:448-452.
- Pfaller MA, Herwaldt LA: **Laboratory, clinical, and epidemiological aspects of coagulase-negative staphylococci.** *Clin Microbiol Rev* 1998, **1**(3):281-299.
- Heikens E, Fleer A, Paauw A, Florijn A, Fluit AC: **Comparison of genotypic and phenotypic methods for species-level identification of clinical isolates of coagulase-negative staphylococci.** *J Clin Microbiol* 2005, **43**:2286-2290.
- Weinstein MP, Mirrett S, Van Pelt L, McKinnon M, Zimmer BL, Kloos W, Reller LB: **Clinical importance of identifying coagulase-negative staphylococci isolated from blood cultures: Evaluation of MicroScan Rapid and Dried Overnight Gram-Positive Panels versus a conventional reference method.** *J Clin Microbiol* 1998, **36**:2089-2092.
- Becker K, Harmsen D, Mellmann A, Meier C, Schumann P, Peters G, von Eiff C: **Development and evaluation of quality-controlled ribosomal sequence database for 16S ribosomal DNA-based identification of Staphylococcus species.** *J Clin Microbiol* 2004, **42**:4988-4995.
- Bosshard PP, Zbinden R, Abels S, Böttlinghaus B, Altwegg M, Böttger EC: **16S rRNA gene sequencing versus the API 20 NE System and the VITEK 2 ID-GNB Card for identification of nonfermenting gram-negative bacteria in the clinical laboratory.** *J Clin Microbiol* 2006, **44**:1359-1366.
- von Eiff C, Peters G, Becker K: **The small colony variant (SCV) concept – the role of staphylococcal SCVs in persistent infections.** *Injury* 2006, **37**:S26-S33.
- von Eiff C, Vaudaux P, Kahl BC, Lew D, Emler S, Schmidt A, Peters G, Proctor RA: **Bloodstream infections caused by small-colony variants of coagulase-negative staphylococci following pacemaker implantation.** *Clin Infect Dis* 1999, **29**:932-4.
- Clarridge JE 3rd: **Impact of 16S rRNA gene sequence analysis for identification of bacteria on clinical microbiology and infectious disease.** *Clin Microbiol Rev* 2004, **17**:840-862.
- Woo PC, Ng KH, Lau SK, Yip KT, Fung AM, Leung KW, Tam DM, Que TL, Yuen KY: **Usefulness of MicroSeq 500 16S ribosomal DNA-based bacterial identification system for identification of clinically significant bacterial isolates with ambiguous biochemical profiles.** *J Clin Microbiol* 2003, **41**:1996-2001.
- Fontana C, Favaro M, Pelliccioni M, Pistoia ES, Favalli C: **Use of the MicroSeq 500 16S rRNA gene-based sequencing for Identification of bacterial isolates that commercial automated systems failed to identify correctly.** *J Clin Microbiol* 2005, **43**:615-619.
- Carretto E, Barbarini D, Couto I, De Vitis D, Marone P, Verhoef J, De Lencastre H, Brisse S: **Identification of coagulase-negative staphylococci other than Staphylococcus epidermidis by automated ribotyping.** *Clin Microbiol Infect* 2005, **11**:177-184.
- Layer F, Ghebremedhin B, Moder KA, König W, König B: **Comparative study using various methods for identification of Staphylococcus species in clinical specimens.** *J Clin Microbiol* 2006, **44**:2824-2830.
- Aldea-Mansilla C, García de Viedma D, Cercenado E, Martín-Rabadán P, Marín M, Bouza E: **Comparison of phenotypic with genotypic procedures for confirmation of coagulase-negative staphylococcus catheter-related bloodstream infections.** *J Clin Microbiol* 2006, **44**:3529-3532.
- Hall L, Doerr KA, Wohlfiel SL, Roberts GD: **Evaluation of the MicroSeq System for identification of Mycobacteria by 16S ribosomal DNA sequencing and its integration into a routine clinical Mycobacteriology laboratory.** *J Clin Microbiol* 2003, **41**:1447-1453.
- Patel JB, Leonard DG, Pan X, Musser JM, Berman RE, Nachamkin I: **Sequence-based identification of Mycobacterium species using the MicroSeq 500 16S rDNA Bacterial Identification System.** *J Clin Microbiol* 2000, **38**:246-251.
- Tang YW, Ellis NM, Hopkins MK, Smith DH, Dodge DE, Persing DH: **Comparison of phenotypic and genotypic techniques for identification of unusual aerobic pathogenic gram-negative bacilli.** *J Clin Microbiol* 1998, **36**:3674-3679.
- Cloud JL, Conville PS, Croft A, Harmsen D, Witebsky FG, Carroll KC: **Evaluation of Partial 16S ribosomal DNA sequencing for identification of Nocardia species by using the MicroSeq 500 System with an expanded database.** *J Clin Microbiol* 2004, **42**:578-584.
- Hussain Z, Stoakes L, Stevens DL, Schieven BC, Lannigan R, Jones C: **Comparison of the MicroScan System with the API Staph-Ident System for species identification of coagulase-negative staphylococci.** *J Clin Microbiol* 1986, **23**:126-128.
- Skulnick M, Patel MP, Low DE: **Evaluation of five commercial systems for identification of coagulase-negative staphylococci to species level.** *Eur J Clin Microbiol Infect Dis* 1989, **8**:1001-1003.
- Stoakes L, Schieven BC, Ofori E, Ewan P, Lannigan R, Hussain Z: **Evaluation of MicroScan Rapid Pos Combo Panels for identification of staphylococci.** *J Clin Microbiol* 1992, **30**:93-95.
- Ben-Ami R, Navon-Venezia S, Schwartz D, Schlezinger Y, Mekuzas Y, Carmeli Y: **Erroneous reporting of coagulase-negative staphylococci as Kocuria spp. by Vitek 2 System.** *J Clin Microbiol* 2005, **43**:1448-1450.
- Ligozzi M, Bernini C, Bonora MG, De Fatima M, Zuliani J, Fontana R: **Evaluation of the VITEK 2 System for identification and antimicrobial susceptibility testing of medically relevant gram-positive cocci.** *J Clin Microbiol* 2002, **40**:1681-1686.
- Eigner U, Schmid A, Wild U, Bertsch D, Fahr A-M: **Analysis of the Comparative Workflow and Performance Characteristics of the VITEK 2 and Phoenix Systems.** *J Clin Microbiol* 2005, **43**:3829-3834.
- Delmas J, Chacornac JP, Robin F, Giammarinaro P, Talon R, Bonnet R: **Evaluation of the Vitek 2 System with a Variety of Staphylococcus Species.** *J Clin Microbiol* 2008, **46**:311-313.
- Kim BJ, Park KY, Kim JY, Ahn JY, Won CH, Lee JH, Rho NH, Kim SH, Cho SY, Kwon OS, Huh CH, Youn SW, Kim MN, Ro BY: **Comparative study of benzoyl peroxide versus clindamycin phosphate in treatment of pitted keratolysis.** *Korean J Med Mycol* 2005, **10**:144-150.

Publish with **BioMed Central** and every scientist can read your work free of charge

"BioMed Central will be the most significant development for disseminating the results of biomedical research in our lifetime."

Sir Paul Nurse, Cancer Research UK

Your research papers will be:

- available free of charge to the entire biomedical community
- peer reviewed and published immediately upon acceptance
- cited in PubMed and archived on PubMed Central
- yours — you keep the copyright

Submit your manuscript here:  
http://www.biomedcentral.com/info/publishing\_adv.asp

